



# Certificate of Compliance

Certificate: 2494608

Master Contract: 163467

Project: 2494608

Date Issued: April 10, 2012

Issued to: Spellman High Voltage Electronics

Corporation  
475 Wireless Blvd  
Hauppauge, NY 11788  
USA  
Attention: Godrej Mehta

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*



*Daniel Peressotti*

Issued by: Daniel Peressotti

## PRODUCTS

**CLASS 8780 01** - MEDICAL ELECTRICAL EQUIPMENT/SYSTEMS

**CLASS 8780 81** - MEDICAL ELECTRICAL EQUIPMENT/SYSTEMS - Certified to US Standards

Power Distribution Unit (P/N 406668-004) with DC-DC Converter (P/N 406550-001), Model PDUX3685, permanently installed, 3 phase; rated:

Input: 380-480 Vac, 3 phase, 50/60 Hz, 250 A Output: 235 Vac, 50/60 Hz, 40 A (Console Power); 235 Vac, 50/60 Hz, 40 A (Gantry Rotor); 235 Vac, 50/60 Hz, 40 A (Gantry Stator); 480 Vac, 3 phase, 50/60 Hz, 40 A (Gantry Motion); 480 Vac, 3 phase, 50/60/Hz, 15 A (Compressor); +/-375Vdc, 132 kW (from DC-DC)

1. Type of protection against electric shock: Class I
2. Degree of protection against electric shock: No applied part
3. Degree of protection against ingress of water: IPX0